



**CITY OF ST. LOUIS
DEPARTMENT OF FINANCE
OFFICE OF THE SUPPLY COMMISSIONER**

**CAROL L SHEPARD, CPA
SUPPLY COMMISSIONER**

**FRANCIS G. SLAY
MAYOR**

**1200 MARKET ST RM 324
ST LOUIS MO 63103
PHONE: 314-622-4580
FAX: 314-622-4141**

DATE: MARCH 2, 2015

ADDENDUM #1 - QUOTE: #41515Q1122 – THERMO SCIENTIFIC GALLERY

BID OPENING DATE: 3/11/15 @ 12:00 NOON

TO: All Bidders

FROM: Stephen M. Gregali – gregalis@stlouis-mo.gov

FOR: WATER DEPARTMENT – HOWARD BEND WATER TREATMENT PLANT

NOTE: TECHNICAL SPECIFICATIONS PAGE #4 OF 4

CORRECTED LINE - FIRST BULLET POINT: Please make note of the new part number. Sheet attached.

- **Gallery Plus ECM w/Water, part #986210011**

ADDENDUM RECEIVED AND ACKNOWLEDGED

Name of Firm

Representative

Signature

Date

Questions: Contact Stephen M. Gregali: gregalis@stlouis-mo.gov

City of St. Louis
Public Utilities
Water Division
Supply & Purifying Section

Req. No. 1122
Date: 2/2/15
Page 4 of 4

TECHNICAL SPECIFICATIONS FOR:
Thermo Scientific Gallery Plus with EMC w/Water

ITEM & DESCRIPTION

BIDDER'S DEVIATIONS

The system shall include the following components:

- Gallery Plus - ECM w/Water, part # ~~968110011~~ **986210011**
- PC Accessories & 2D barcode Scanner Included, part # 9860121
- Gallery Start-up kit, part # 986100
- ECM Start-up kit, part # 986200
- Installation and training

The system shall include all necessary reagents and electrodes to analyze:
Alkalinity, Ammonia, Calcium, Conductivity, Fluoride, Hardness (Total),
Magnesium, and pH.

All methods must reference either EPA or Standard Methods for the
Examination of Water and Wastewater.

The system shall come with a one year warranty covering parts and labor.